*Document reference no.: 220715/1*
*Report no.: RN:-0513*

# EU DECLARATION OF CONFORMITY

*This certifies that the following designated products:*

| | |
|---|---|
| **HI9810-61** | Portable pH,EC,TDS Meter |
| **HI9813-61** | Portable pH,EC,TDS and Temperature Meter with CAL Check |
| **HI1285-61** | pH, EC, TDS and Temperature Multiparameter Probe |

| | | |
|---|---|---|
| Are in compliance with: | Directive 2011/65/EU | RoHS2 |
| | Directive 2014/30/EU | EMC and its standards |
| | Directive 2014/35/EU | Low voltage and its standards |

*The instruments are in conformity with the following harmonized standards used or references to the other technical specification:*

| | | |
|---|---|---|
| EN 63000:2018 | Restriction of the use of certain hazardous substances (RoHS) | |
| EC Regulation 1907/2006 | Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) | |
| EN 61326-1: 2013 | Electrical Equipment of Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards: | Class A |
| EMC requirements Part 1: | General requirements and its standards: | |
| EN 61000-4-2: 2009, EMC | Electrostatic Discharge Immunity Test | 4kV/8kV (contact / air) |
| EN 61000-4-3: 2020, EMC | Radiated RF, Electromagnetic Field Immunity Test | 6V/m (80MHz to 1GHz) |

*The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.*

This declaration is the responsibility of the manufacturer:

**S.C. Hanna Instruments S.R.L.**
Str. Hanna Nr.1, loc Nușfalău, jud. Sălaj,
cod 457260, ROMANIA

**Hanna Instruments, Inc.**
Highland Industrial Park, 584 Park East Drive,
Woonsocket, RI 02895, USA
Phone: +1 401 7657500

CE

Test Laboratory

*This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.*
*The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.*

Hanna Instruments SRL

Test performed by: Ervin Jozsa / ETO Eng. /
[ Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /
[ Name / Title of Signatory / Signature]

Issue place / Date: Nușfalău / February 15, 2022
[Place / Date in format: Month, dd, yyyy]