Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**pH 10.01 Calibration Buffer, Blue**
Catalog Number:	P1005
Lot Number:	P10052542601
Manufacture Date:	09/11/26
Expiration Date:	09/11/28

Property	Specification	Result
Appearance	Clear, blue liquid	Pass
pH at 25 °C	10.01 ± 0.02	10.01

Conforms to RMR Specification QA-PD-P1005 Revision B.

Prepared with reagent grade materials and high purity water. Measurement results are certified traceable to NIST.

Certificate approved on 09/17/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.