Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**pH 4.01 Calibration Buffer, Red**
Catalog Number:	P1001
Lot Number:	P10012512601
Manufacture Date:	09/08/26
Expiration Date:	09/08/28

Property	Specification	Result
Appearance	Clear, red liquid	Pass
pH at 25 °C	4.01 ± 0.01	4.01

Conforms to RMR Specification QA-PD-P1001 Revision B.

Prepared with reagent grade materials and high purity water. Measurement results are certified traceable to NIST.

Certificate approved on 09/11/26 by:



Emily Wells
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.