Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS


Product:	**pH 7.00 Calibration Buffer, Yellow**
Catalog Number:	P1002
Lot Number:	P10022112601
Manufacture Date:	07/30/26
Expiration Date:	07/30/28

Property	Specification	Result
Appearance	Clear, yellow liquid	Pass
pH at 25 °C	7.00 ± 0.01	7.00

Conforms to RMR Specification QA-PD-P1002 Revision A.

Prepared with reagent grade materials and high purity water. Measurement results are certified traceable to NIST.

Certificate approved on 08/06/26 by:



Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.